Powered by Mitti

AED Monthly Checklist

23 Oct 2026

Complete

Score	15 / 15 (100%)	Flagged items	0	Actions	0

Location of AED	Oak View, Main Street, Fenton, Newark England NG23 5DE United Kingdom (53.04738833158851, -0.6886448542712633)
AED Serial No.	0000137073
Checked on	23 Oct 2026 10:47 BST
Checked by	Graham Doney

AED Monthly Checklist	15 / 15 (100%)
Is the AED secure in its container?	Yes
Check AED is on standby mode - green flashing light	☑

If AED is not on standby mode, verify AED operation

Verify AED Operation	4 / 4 (100%)
Open AED lid / Turn on AED	☑
Wait for status indicator to change from RED to GREEN after five seconds	☑
Close AED lid / Turn off AED	☑
Wait for status indicator to change from RED to GREEN after five seconds	☑
Are the electrode pads in the AED container?	Yes
Check expiration date of electrode pads	☑
Electrode Pads Expiration Date:	24 Dec 2026
Is there an extra set of electrode pads?	Yes
Check expiration date of AED battery	☑

The AED battery is expired if 4 years have passed since its installation.

AED Battery Installation Date:	10 Jul 2023

Check PPE/Ready Kit	4 / 4 (100%)
Mask	☑
Scissors	☑
Razor	☑
Gloves	☑
Test AED container tamper alarm	☑

IMPORTANT Maintenance	
If electrode pads were expired, did you replace them?	N/A
Input the date the electrode pads were replaced:	1 Jun 2025
If AED battery was expired, did you replace it?	N/A
Input the date the AED battery was replaced:	10 Jul 2023
If the AED container tamper alarm was not working, did you change its battery?	N/A

Signature



23 Oct 2026 10:53 BST