



AED Monthly Checklist

28 Dec 2025

Complete

Score	15 / 15 (100%)	Flagged items	0	Actions	0
-------	----------------	---------------	---	---------	---

Location of AED	Manor View, Main St, Newark NG23 5DE, UK (53.0473157, -0.6886782)
AED Serial No.	0000137073
Checked on	28 Dec 2025 09:05 GMT
Checked by	Graham Doney

AED Monthly Checklist

15 / 15 (100%)

Is the AED secure in its container?**Check AED is on standby mode - green flashing light**

If AED is not on standby mode, verify AED operation

Verify AED Operation

4 / 4 (100%)

Open AED lid / Turn on AED**Wait for status indicator to change from RED to GREEN after five seconds****Close AED lid / Turn off AED****Wait for status indicator to change from RED to GREEN after five seconds****Are the electrode pads in the AED container?****Check expiration date of electrode pads****Electrode Pads Expiration Date:**

24 Dec 2026

Is there an extra set of electrode pads?**Check expiration date of AED battery**

The AED battery is expired if 4 years have passed since its installation.

AED Battery Installation Date:

10 Jul 2023

Check PPE/Ready Kit

4 / 4 (100%)

Mask**Scissors****Razor****Gloves****Test AED container tamper alarm**

IMPORTANT Maintenance

If electrode pads were expired, did you replace them?

N/A

Input the date the electrode pads were replaced:

1 Jun 2025

If AED battery was expired, did you replace it?

N/A

Input the date the AED battery was replaced:

10 Jul 2023

If the AED container tamper alarm was not working, did you change its battery?

N/A

Signature



Graham E Doney
28 Dec 2025 09:22 GMT